

Document downloaded from:

<http://hdl.handle.net/10251/43198>

This paper must be cited as:

Trujillo Guillen, M.; Castellvi, Q.; Burdío, F.; Sánchez Velázquez, P.; Ivorra, A.; Andaluz, A.; Berjano, E. (2013). Can electroporation previous to radiofrequency hepatic ablation enlarge thermal lesion size? A feasibility study based on theoretical modelling and in vivo experiments. *International Journal of Hyperthermia*. 29(3):211-218.
doi:10.3109/02656736.2013.777854.



The final publication is available at

<http://dx.doi.org/10.3109/02656736.2013.777854>

Copyright Informa Healthcare

**Can electroporation previous to radiofrequency hepatic ablation
enlarge thermal lesion size? Feasibility study based on theoretical
modeling and in vivo experiments**

MACARENA TRUJILLO¹, QUIM CASTELLVÍ², FERNANDO BURDÍO³,
PATRICIA SÁNCHEZ VELAZQUEZ³, ANTONI IVORRA², ANNA ANDALUZ⁴ &
ENRIQUE BERJANO⁵

*¹Instituto Universitario de Matemática Pura y Aplicada, Universitat Politècnica de
València, Spain*

*² Departament de Tecnologies de la Informació i la Comunicació, Universitat Pompeu
Fabra, Barcelona, Spain*

³ General Surgery Department, Hospital del Mar, Barcelona, Spain

*⁴ Departament de Medicina i Cirurgia Animals, Facultat de Veterinària, Universitat
Autònoma de Barcelona, Barcelona, Spain*

*⁵ Biomedical Synergy, Electronic Engineering Department, Universitat Politècnica de
València, Spain*

*To whom all correspondence should be addressed:

Dr. Enrique Berjano, Biomedical Synergy, Electronic Engineering Department (7F)

Universitat Politècnica de València, Camino de Vera, 46022, Valencia, Spain

Phone: 34–963877607, Fax: 34–963877609, Email: eberjano@eln.upv.es

Abstract

Purpose: To assess the feasibility of a hybrid ablative technique based on applying electroporation (EP) pulses just before conducting radiofrequency ablation (RFA). The rationale was that the EP-induced reduction in blood perfusion could be sufficient to reduce the thermal sink effect and hence to increase the coagulation volume in comparison to that created exclusively by RFA.

Materials and methods: Modeling study and *in vivo* experimental study. A Cool-tip RF applicator was used both for EP and RFA.

Results: Overall, the results did not show any synergy effect from using the hybrid technique. Applying EP pulses prior to RFA did not increase the coagulation zone obtained and the lesions were almost identical. Additional computer simulations provided an explanation for this; the effect of reducing blood perfusion by thermal damage during RFA completely masks the effect of reducing blood perfusion by EP. This is because both thermal damage and EP affect the same zone, i.e. the tissue around the electrode.

Conclusions: Our computer modeling and *in vivo* experimental findings suggest that the combination of EP and RFA with monopolar applicators does not provide an additional benefit over the use of RFA alone.

Keywords: *Electroporation, radiofrequency ablation*

Running head: Electroporation before RF ablation

Introduction

Radiofrequency (RF) ablation (RFA) is a minimally invasive ablative technique used to treat certain tumors [1, 2] by using RF current (≈ 500 kHz) to produce a coagulation zone characterized by a thermal lesion which destroys the target tissue by necrosis. The coagulation zone is defined as the zone containing irreversibly damaged tissue, i.e. with temperatures above 50°C [3]. In order to treat large tumors, therefore, the RF applicator should be able to create large coagulation zones. One of the most important factors limiting coagulation zone size is blood perfusion through the capillaries, which causes a thermal sink effect [4]. In fact, the coagulation zone can be increased during RFA when blood flow is reduced by clamping the organ's main artery [5]. Unfortunately, this maneuver produces a generalized ischemia in the organ and hence is not a satisfactory solution.

In Electroporation (EP) cell membrane permeability to ions and macromolecules is increased after exposing the cell to high electric field short pulses. According to the number of pulses, their magnitude, duration and other factors, the permeability will be either temporary and will not compromise cell viability (i.e. "reversible electroporation") or will be permanent, or too intense, so that cell homeostasis will be severely disrupted and the cell will end up dying by necrotic or apoptotic processes (i.e. "irreversible electroporation"). Both reversible and irreversible EP modes have important applications in biotechnology and medicine. For instance, reversible EP is now a routine technique in microbiology labs for *in vitro* gene transfection. During the last fifteen years, reversible electroporation has been applied experimentally and clinically to living tissues for *in vivo* gene therapy (electrogenetherapy) and for enhancing the penetration of anti-cancer drugs into cancer cells (electrochemotherapy) [6]. More recently, irreversible EP has been proposed and demonstrated as a minimally

invasive tissue ablation method [7].

An interesting outcome of performing EP in living tissues is the immediate local reduction in blood flow. Blood perfusion has been reduced by about 70% for example in subcutaneous tumors a few seconds after receiving eight 100 μ s pulses at a magnitude of 130000 V/m and a repetition rate of 1 Hz [8]. This reduction is related to microvascular occlusion of capillaries and arterioles and in some cases reverts in a matter of minutes or hours [9, 10]. Two main explanations have been hypothesized for this phenomenon: 1) a sympathetically mediated constriction of smooth vascular musculature would explain the first rapid phase of blood flow reduction and 2) damage to the endothelial monolayer by electroporation would cause liquid extravasation, which would increase parenchyma pressure, collapse vessels and thus cause blood occlusion [8].

With these results in mind, we hypothesized that the reduced blood perfusion caused by EP pulses applied prior to RFA could be sufficient to reduce the thermal sink effect, and hence to increase coagulation volume more than that created solely by RFA. In order to assess the feasibility of this hybrid ablative technique, we first planned a theoretical modeling study in which a Cool-tip RF applicator was used to apply both EP pulses and RF power to hepatic tissue. We then conducted an experimental study based on an *in vivo* model to validate the theoretical results obtained.

Materials and methods

Theoretical modeling: General characteristics

The theoretical model was based on a coupled EP+RF problem, each part of which was, in turn, based on a coupled electric-thermal problem and solved numerically using the Finite Element Method (FEM) by means of COMSOL Multiphysics 3.5a (COMSOL,

Burlington, MA, USA). The geometry and dimensions of the theoretical model are shown in Fig. 1. The problem presented axial symmetry, so that a 2D analysis could be conducted. The model included three different materials: plastic partially covering the electrode, a metallic electrode, and hepatic tissue. The values of the characteristics of these materials are given in Table I [11-14]. It should be noted that the electrical conductivity value for hepatic tissue corresponds to the RF problem; the conductivity value for the EP problem is somewhat lower (see the following subsection). The dispersive electrode was modeled as an electrical condition on boundaries at a distance from the active electrode.

Theoretical model: EP problem

As described below in the RF problem, the electric field in the tissue due to an EP pulse was computed by means of the FEM tool. The *Quasi-Statics/Electric* application mode of the *AC/DC Module* of COMSOL was employed to obtain the electric field when a pulse of 2000 V was applied between the active electrode and dispersive electrode (Fig. 1).

Non-linear resistive behavior was assumed for the hepatic tissue (Gompertz sigmoid expression, Eq. 1) to model the increase in tissue electrical conductivity during pulse delivery due to EP [15]. It has recently been shown that modeling this effect significantly improves the accuracy of these simulations [16,17]. DC conductivity at low electric field magnitudes ($\sigma_0 = 0.08$ S/m) and DC conductivity at very high fields ($\sigma_{\text{MAX}} = 0.31$ S/m = $\sigma_0 + \Delta\sigma$) were obtained from [15]. The sigmoid behavior and the parameters that determine its shape were adjusted from data in [18]. Gompertz sigmoid function was chosen instead of other possible sigmoid functions, such as the sigmoid

expression employed in [18], because its shape parameters are easily identifiable in the expression.

$$\sigma = \sigma_0 + \Delta\sigma e^{-e^{-0.01 \frac{(|\mathbf{E}|-50000)}{80}}} \quad (1)$$

The simulated time was not relevant, since no time-dependant properties or effects were modeled. Nevertheless, for the sake of conceptual consistency, the electric field for computing perfusion was recorded at the end of a pulse of 100 μs . This pulse included a rise ramp of 5 μs to facilitate the convergence of the solution required by the iterative process, in which the electric field distribution depends on the conductivity of the previous step and vice versa.

Prior to the RF problem, the computed electric field magnitude at the end of the 100 μs was scaled according to the following expression (2) to obtain the reduction in blood flow:

$$\alpha = \begin{cases} 1 & 0 < |\mathbf{E}| \leq 75000 \text{ V/m} \\ 1 - 0.7 \frac{|\mathbf{E}| - 75000}{130000 - 75000} & 75000 < |\mathbf{E}| \leq 130000 \text{ V/m} \\ 0.3 & |\mathbf{E}| > 130000 \text{ V/m} \end{cases} \quad (2)$$

where α represents the relative blood perfusion level; 1 indicates the normal perfusion level (i.e. before EP or in tissue zones in which the electric field magnitude is insufficient to cause any blood flow disturbance) and 0 represents the complete disruption of blood flow. Note that a linear decrease in blood perfusion was modeled between two electric field values (75000 V/m and 130000 V/m) and that the minimum relative blood perfusion level achievable by EP is 0.3.

The above blood perfusion expression was obtained by approximation of data in [8], which corresponds to experimental measurements in subcutaneous tumors after a

sequence of eight 100 μ s pulses. These original data together with the function used to model them (α) are depicted in Figure 2.

Quantitative liver data is not presently available, but as similar quantitative data was obtained for skeletal muscle [19] and qualitative observations on liver also indicate a drastic blood perfusion reduction in equivalent conditions [9], this suggests that the above-described model is sufficiently accurate for the purposes of this study.

Theoretical model: RF problem

The governing equation for the thermal problem was the Pennes Bioheat equation [20]. To incorporate the phase change [21-22] into this equation we used the enthalpy method [23]:

$$\frac{\partial(\rho h)}{\partial t} = \nabla(k\nabla T) + q + Q_m + Q_p \quad (3)$$

where ρ was density of tissue, h enthalpy, T temperature, t time, k thermal conductivity, q heat source produced by RF power, Q_m the metabolic heat generation (not considered in RF ablation) and Q_p was heat loss from blood perfusion.

For biological tissues, enthalpy is related to tissue temperature by the following expression [24]:

$$\rho h = \begin{cases} \rho_l c_l T & 0 < T \leq 99^\circ C \\ \rho h(99) + h_{fg} C \frac{(T - 99)}{(100 - 99)} & 99 < T \leq 100^\circ C \\ \rho h(100) + \rho_g c_g (T - 100) & T > 100^\circ C \end{cases} \quad (4)$$

where ρ_i and c_i were tissue density and specific heat of normal tissue ($i=l$) and the post-phase-change tissue ($i=g$), respectively, h_{fg} was the product of water latent heat of vaporization and water density at 100°C and C was tissue water content inside the liver (68%) [25].

The heat source by RF power q (Joule losses) was given by $q=\mathbf{J}\mathbf{E}$, where \mathbf{J} is current density and \mathbf{E} is electric field strength. Both were obtained from the electrical problem, in which we used the Laplace equation $\nabla^2 V=0$ as governing equation, V being the voltage. The electric field was calculated by means of $\mathbf{E} = -\nabla V$ and \mathbf{J} using Ohm's law ($\mathbf{J}=\sigma\mathbf{E}$), where σ is the electrical conductivity. We used a quasi-static approach, i.e. the tissues were considered as purely resistive [26] due to the value of the frequencies used in RF (≈ 500 kHz) and for the geometric area of interest (electrical power is deposited in a very small zone close the electrode). We considered temperature-dependent functions for the electrical and thermal conductivity, σ and k [26]: σ grew exponentially $+1.5\%/^{\circ}\text{C}$ up to 100°C [21], between 99 and 100°C was kept constant and then σ decreased linearly 4 orders for five degrees (i.e. up to 105°C) [23,27]. For k we considered linear growth of $0.0015/^{\circ}\text{C}$ up to 100°C , after which k was kept constant [23]. The term Q_p was obtained from:

$$Q_p = \alpha \cdot \beta \cdot \rho_b \cdot c_b \cdot \omega_b \cdot (T_b - T) \quad (5)$$

where ρ_b was density of blood (1000 kg/m^3) [28], c_b specific heat of blood ($4180 \text{ J/Kg}\cdot\text{K}$) [27], T_b blood temperature (37°C), ω_b blood perfusion coefficient ($6.4 \times 10^{-3} \text{ s}^{-1}$) [28], α relative blood perfusion which varied between 0 and 1 obtained in the EP problem, and β was a coefficient which took the values of 0 and 1, depending on the value of the local thermal damage Ω : $\beta = 0$ for $\Omega \geq 1$, and $\beta = 1$ for $\Omega < 1$ [29]. Ω was assessed by the Arrhenius damage model [30], which associates temperature with exposure time using a first-order kinetics relationship:

$$\Omega(t) = \int_0^t A e^{\frac{-\Delta E}{RT}} dt \quad (6)$$

where R was the universal gas constant, A was a frequency factor and ΔE was the activation energy for the irreversible damage reaction. The parameters A and ΔE were dependent on tissue type; for liver we used those proposed by Jacques et al [31]: $A = 7.39 \times 10^{39} \text{ s}^{-1}$ and $\Delta E = 2.577 \times 10^5 \text{ J/mol}$. We employed two thermal damage contours to compute the lesion dimension contours, which correspond to a damage integral (Ω) of 1 and 4.6 (63% and 99% probability of cell death respectively) [29].

Electrical boundary conditions were: zero current density in the transverse direction to the symmetry axis and inside the electrode. Zero voltage was set in the dispersive electrode. A voltage of 80 V was applied in the active electrode. Thermal boundary conditions were: Null thermal flux in the transversal direction to the symmetry axis and constant temperature of 37°C in the dispersive electrode. Initial temperature of the tissue was considered to be 37°C. The liquid circulating inside the electrode produced a cooling effect modeled by means of a thermal convection coefficient h with a value of 6,668 W/K·m² and a coolant temperature of 10°C. The value of h was calculated by considering a length of 30 mm and a flow rate of 45 mL/min through an area of $3.93 \times 10^{-7} \text{ m}^2$, which is equivalent to half of the cross section of the inner diameter of the electrode (see Fig. 1).

The mesh used was heterogeneous, with a finer mesh size at the electrode-tissue interface, where the highest gradient was expected. The size of the finer mesh was estimated by a convergence test. We used the value of the maximum temperature (T_{\max}) reached in the liver at roll-off time (at which tissue impedance was 30 Ω higher than the initial [13]) as a control parameter in these analyses. When there was a difference of less than 0.5% in the control parameter between simulations, we considered the former

mesh size as appropriate. Analogously, we used a value of 1 for the time-step obtained from a similar convergence test scheme.

A pulsed RF protocol was used for energy delivery, consisting of the application of 80 V up to roll-off time, after which energy was not applied for 15 s and then 80 V was again applied until the next roll-off. This protocol was followed for 12 minutes and implemented by the connection between COMSOL and MATLAB.

Experimental setting

A total of 12 procedures (one in each hepatic lobe) were performed on the liver of three female Landrace pig under general anesthesia through midline laparotomy. The procedure was approved by the Ethical Commission of the Universitat Autònoma de Barcelona (Authorization Number DAAM 6267 and CEEAH 1256). In each animal, in the first lobe RFA only was performed, in the second an EP protocol only and in the third and fourth the EP+RF procedure was applied. RFA was conducted immediately after EP. In both EP and RFA a single Cool-tip applicator with 3 cm-long electrode (Valleylab, Boulder, CO, USA) was applied parallel to the suprahepatic main vein of each lobe.

EP pulses were applied in a monopolar arrangement between the Cool-tip applicator electrode and a single dispersive electrode positioned on the interscapular area. A custom-made generator was used to deliver 8 square-wave pulses with duration of 100 μ s and a magnitude of 2000 V at a repetition frequency of 1 Hz. A high voltage probe (P5100, Tektronix) and a clamp-on current probe (A622, Tektronix) were used in combination with a digital oscilloscope (DS01014A, Agilent) to measure voltage and current signals respectively, to ensure proper application of the pulses.

RFA was also conducted in a monopolar mode between the applicator and the

dispersive electrode, but in this case the electrode was internally cooled by a Watson Marlow peristaltic pump (Wilmington, MA, USA), which delivered 0°C saline at 20 mL/min, beginning 2 minutes prior to RF power deposition. RF power was applied with a CC-3 generator (Radionics, Burlington, MA, USA) using a pulsed algorithm based on an impedance control mode for 12 minutes, as employed in clinical use. In each animal, when the four ablations were complete the animal was euthanized using an IV dose of sodium pentobarbital. In any case more than one hour elapsed from any thermal injury to euthanasia.

We immediately conducted a macroscopic assessment of the coagulations as follows: the liver was cut into 1 cm slices perpendicular to the applicator axis in order to measure macroscopically maximum and minimum diameters of the coagulation zone. All the measurements were based on the central “white zone” induced by thermal therapies, avoiding the red area of hyperemia [32]. Mean values of maximum and minimum diameters were compared through the groups using the U-Mann-Whitney’s test. Differences in variables were considered to be significant at a threshold of $p < 0.05$. Statistical analyses were performed with statistical software (version 19.0, SPSS, Chicago, IL, USA).

Results

Numerical results

As a result of the EP pulse application the tissue perfusion changed according to the relative blood perfusion level (α). Figure 3 shows the electric field distribution joint with the resulting blood perfusion at the end of the EP pulse. In the RF problem we first simulated two cases and then compared their results: the EP+RF procedure (Case 1) and RFA alone (Case 2). Fig. 4A and 4B show voltage and impedance evolution for the

Case 1. The EP pulse application interval (7 seconds) is insignificant compared to the 12-minute interval in RFA, and hence the EP period is not plotted. The first roll-off (sudden impedance rise) was reached at 60 s. The subsequent roll-offs were observed 30 s later and the time between roll-offs was decreasing linearly to the fourth roll-off when the subsequent roll-offs were reached more or less periodically with an interval of ≈ 23 s. Note that after every roll-off there was a 15 s period in which almost no voltage was applied. Fig. 5 shows the 2D temperature distributions and the thermal damage contour ($\Omega=1$ and $\Omega=4.6$, black lines) at the end of RFA (12 minutes) in Case 1. In Table II are shown the transverse diameter of the thermal lesion for both thermal damage contours. The maximum temperature reached in the tissue was 110°C. In Figure 6 we observed the temperature distribution in case 1 at the end of the RF application (12 minutes) and the contour lines corresponding to the blood perfusion level (α) obtained from the EP problem.

The theoretical results from RFA-alone (Case 2) were extremely similar to the EP+RF case. The first roll-off occurred at 61 s, and the following roll-offs were observed with a similar periodicity than in case 1. Fig. 4C shows the impedance evolution for the RFA-alone case. Note that in Figs. 4B and 4C the roll-off episodes occur almost at the same time. Likewise, the temperature distributions (not shown) were almost identical to the EP+RF case, with similar values for the minor diameter of the lesion (see Table II).

Since the computer results from both cases were almost identical, we planned additional computer simulations in order to provide a possible explanation. We hypothesized that the dependence of blood perfusion on thermal damage (term β in Eq. 5) could drastically reduce the impact of EP on the perfusion term (α in Eq. 5), specially

when the EP effect is confined exclusively to the tissue adjacent to the electrode. To confirm this hypothesis we simulated two additional cases: EP+RF with the blood perfusion term independent of thermal damage (Case 3) and also RFA alone with perfusion independent of thermal damage (Case 4). The results from these two additional cases were found to be different to Cases 1 and 2. While the first roll-offs occurred at 68 s and 74 s in Cases 3 and 4, respectively, the interval between roll-off episodes was also different: between 25 and 35 s in Case 3 and between 25 s and 40 s in Case 4, in which roll-offs were less frequent. These longer differences imply fewer roll-off episodes in 12 minutes. The transverse diameters of the coagulation zone for $\Omega=1$, were 2.84 cm and 2.50 cm for the Cases 3 and 4 respectively.

The case with Q_p term set to zero would correspond with a RF ablation with total cessation in blood flow such as that found during Pringle maneuver. We conducted this additional simulation and the results showed a diameter of 4.20 cm for $\Omega=1$ and 3.55 cm for $\Omega=4.6$, i.e. larger than in cases 1 and 2, as it is expected.

Experimental results

Figure 7 shows the four lesions created in each hepatic lobe in a representative case. In the first (A), in which EP only was performed, no coagulation “white area” can be observed, but only a two 2-mm wide hyperemia along the electrode path. In the second lobe, in which RFA alone was performed, a “white area” of between 4.8 cm (maximum diameter) and 1.7 mm (minimum diameter) can be seen. In the third and fourth lobes (EP+RF procedure) a similar coagulation “white area” between 4.2 cm (maximum diameter) and 1.6 mm (minimum diameter) were visible. No significant differences

between EP and RF+EP groups were observed either in the maximum or in the minimum diameter (Table II).

Discussion

This feasibility study was undertaken to assess whether the reduced blood perfusion caused by EP pulses applied prior to RFA would significantly reduce the thermal sink effect and hence increase coagulation volume to a greater extent than that created solely by RFA. We first conducted a computer modeling study, followed by a brief experimental study on an *in vivo* model.

Overall, the results suggest that no synergy effect can be achieved from this hybrid protocol, i.e. applying EP pulses prior to RFA does not increase the area of the coagulation zone. The additional computer simulations (Cases 3 and 4) could provide a possible explanation. In the case of RFA alone (Cases 2 and 4), the fact of not considering the effect of thermal damage on blood perfusion (Case 4) is responsible for a reduction of 15% in the transverse diameter as compared to Case 2, in which this factor was taken into consideration (2.94 cm vs. 2.50 cm). This result agrees with other previous findings. However, when comparing the two EP+RFA cases (1 and 3), the transverse diameter decreased by only 4% when blood perfusion was not dependent on thermal damage (2.98 cm vs. 2.84 cm), which suggests that the effect of EP is noticeable in Case 3, i.e. only when the reduced blood perfusion due to thermal damage is not taken into account. In other words, the effect of reducing blood perfusion by thermal damage completely masks the effect of reducing blood perfusion by EP. This is due to the fact that both thermal damage and EP affect the tissue next to the electrode.

On one hand the disagreement between experimental and theoretical minor diameters both in cases EP+RF and RF was due to the different criterion used to assess thermal

damage. Since theoretical modeling used the Arrhenius damage model (with probabilities of cell death of 63% and 99%), it is possible that the computed boundaries do not match very well to the boundaries assessed macroscopically by the “white zone”. On the other hand, in the computation of Ω we used parameters A and ΔE . The values of these parameters are obtained experimentally. Bibliography provides several values for A and ΔE depending on the experiment used to obtain them and they are very different, see for example [33]. We checked that changes in the A value produced great differences in lesion dimensions. Then, we chose A and ΔE values which provided reasonable results, i.e. lesion dimensions in an acceptable range. However, due to the way in which A and ΔE are obtained and the great influence of A in the value of Ω was almost impossible that in our study theoretical predicted lesions were equal to experimental ones. In spite of this disagreement, the objective of the experiments was not to validate the theoretical model, but to confirm the results obtained from computer simulations.

The RF ablation uses a frequency around 500 kHz. Probably over 2 MHz the displacement currents could take a role more and more important, but in the electrosurgical range (300 kHz-1MHz), they can be considered negligible.

Several limitations implicit in this study should be pointed out: 1) the theoretical model only considered the effect of EP on the reduction in blood perfusion, i.e. no late cytophatic effects (apoptosis) were modeled; and, in relation to this, 2) the experimental model did not consider the late effects since the coagulation zones were assessed immediately after the procedure. In spite of these limitations, we think that the effect of EP on reducing blood perfusion, at least with the monopolar arrangement tested here, is confined to the vicinity of the electrode, where the heating is mainly concentrated and the effect on blood perfusion is thus reduced. The impact of the EP pulses in terms of

reducing blood perfusion is therefore masked by the subsequent RFA, which also has the effect of reducing blood flow.

On the other hand, one of the main limitations for thermal therapies is the heat sink by high rate in major blood vessels, which may locally impede the thermal damage to the tumor, permitting neoplastic cell survival near these blood vessels. To date, the vascular effects from EP have not been shown to disrupt blood perfusion through these large vessels. For this reason, the conclusions of our study should be only considered in the case of combined EP+RFA therapy for tumors not adjacent to large vessels.

Conclusion

As far as we know, no previous studies have been conducted to check the possibility of combining RFA and EP to increase the size of the coagulation zones when ablating tumors. Our computer modeling and *in vivo* experimental findings suggest that this combination, when using monopolar applicators, does not provide an additional advantage as compared to the use of RFA alone.

Acknowledgements

This work received financial support from the Spanish “Plan Nacional de I+D+I del Ministerio de Ciencia e Innovación” Grants No. TEC2011-27133-C02-01,02 and TEC2010-17285, from the European Commission through the Marie Curie IRG grant TAMIVIVE (256376) and from the Universitat Politècnica de València (INNOVA11-01-5502; and PAID-06-11 Ref. 1988). The linguistic revision of this paper was funded by the Universitat Politècnica de València, Spain.

Declaration of Interest

The authors report no conflicts of interest.

References

1. Poon RT, Fan ST, Tsang FH, Wong J. Locoregional therapies for hepatocellular carcinoma: a critical review from the surgeon's perspective. *Ann Surg* 2002;235:466–486.
2. Solbiati L, Livraghi T, Goldberg SN, Ierace T, Meloni F, Dellanoce M, et al. Percutaneous radiofrequency ablation of hepatic metastases from colorectal cancer: Long-term results in 117 patients. *Radiology* 2001;221:159–166.
3. Ahmed M, Brace CL, Lee Jr FT, Goldberg SN. Principles of and advances in percutaneous ablation. *Radiology* 2011; 258:351–369.
4. Nikfarjam M, Muralidharan V, Christophi C. Mechanisms of focal heat destruction of liver tumors. *J Surg Res* 2005;127:208–223.
5. Burdio F, Mulier S, Navarro A, Figueras J, Berjano E, Poves I, et al. Influence of approach on outcome in radiofrequency ablation of liver tumors. *Surg Oncol* 2008;17(4):295–299.
6. Marty M, Sersa G, Garbay JR, Gehl J, Collins CG, Snoj M, et al. Electrochemotherapy—An easy, highly effective and safe treatment of cutaneous and subcutaneous metastases: Results of ESOPE (European Standard Operating Procedures of Electrochemotherapy) study. *Eur J Cancer Supplements* 2006;4(11):3–13.
7. Davalos RV, Mir IL, Rubinsky B. Tissue ablation with irreversible electroporation. *Annals Biomed Eng* 2005;33(2):223–231.
8. Sersa G, Cemazar M, Parkins CS, Chaplin DJ. Tumour blood flow changes induced by application of electric pulses. *European Journal of Cancer* 1999;35(4):672–677.
9. Ramirez LH, Orłowski S, An D, Bindoula G, Dzodic R, Ardouin P, et al. Electrochemotherapy on liver tumours in rabbits. *British J Cancer* 1998;77(12):2104–2111.
10. Bellard E, Markelc B, Pelofy S, Le Guerroué F, Sersa G, Teissié J, et al. Intravital microscopy at the single vessel level brings new insights of vascular modification mechanisms induced by electropermeabilization. *J Control Release* 2012; 163(3): 396-403.
11. Berjano EJ, Burdío F, Navarro AC, Burdío JM, Güemes A, Aldana O, et al. Improved perfusion system for bipolar radiofrequency ablation of liver. *Physiol Meas* 2006;27:55–66.
12. Pätz T, Körger T, Preusser T. Simulation of Radiofrequency Ablation Including Water Evaporation. *IFMBE Proceedings* 2009;25:1287–1290.

13. Trujillo M, Alba J, Berjano E. Relation between roll-off occurrence and spatial distribution of dehydrated tissue during RF ablation with cooled electrodes. *Int J Hyperthermia* 2012;28:62–68.
14. Duck F. Physical properties of tissue - A comprehensive reference book. *Academic Press* New York. 1990.
15. Ivorra A, Rubinsky B. In vivo electrical impedance measurements during and after electroporation of rat liver. *Bioelectrochemistry* 2007; (Amsterdam, Netherlands);70(2):287–295.
16. Ivorra A, Mir LM, Rubinsky B. Electric field redistribution due to conductivity changes during tissue electroporation: experiments with a simple vegetal model", *IFMBE Proceedings* 2009;25(13):59–62.
17. Lacković I, Magjarević R, Miklavčič D. Incorporating electroporation-related conductivity changes into models for the calculation of the electric field distribution in tissue. *XII Mediterranean Conference on Medical and Biological Engineering and Computing* 2010:695.
18. Sel D, Cukjati D, Batiuskaite D, Slivnik T, Mir LM, Miklavcic D. Sequential finite element model of tissue electropermeabilization. *IEEE Trans Biomed Eng* 2005;52(5):816–827.
19. Gehl J, Skovsgaard T, Mir LM. Vascular reactions to in vivo electroporation: characterization and consequences for drug and gene delivery. *Biochimica et biophysica acta* 2002;1569(1-3):51–58.
20. Pennes H. Analysis of tissue and arterial blood temperatures in the resting human forearm. *J Appl Physiol* 1948;85:5–34.
21. Pearce J, Panescu D, Thomsen S. Simulation of dioptr changes in radio frequency conductive keratoplasty in the cornea. *WIT Trans Biomed Health* 2005;8:469–477.
22. Yang D, Converse MC, Mahvi DM, Webster JG. Expanding the bioheat equation to include tissue internal water evaporation during heating. *IEEE Trans Biomed Eng* 2007;54:1382–1388.
23. Byeongman J, Aksan A. Prediction of the extent of thermal damage in the cornea during conductive thermokeratoplasty. *J Therm Biol* 2010;35:167–174.
24. Abraham JP, Sparrow EM. A thermal-ablation bioheat model including liquid-to-vapor phase change, pressure- and necrosis-dependent perfusion, and moisture-dependent properties. *Int J Heat Mass Tran* 2007;50:2537–2544.

25. Arata MA, Nisenbaum HL, Clark TW, Soulen MC. Percutaneous radiofrequency ablation of liver tumors with the LeVein probe: is roll-off predictive of response? *J Vasc Interv Radiol* 2001;12:455–458.
26. Berjano EJ. Theoretical modeling for radiofrequency ablation: State of-the-art and Challenges for the future. *Biomed Eng Online* 2006;18;5:24.
27. Tungjitkusolmun S, Woo EJ, Cao H, Tsai JZ, Vorperian VR, Webster JG. Thermal-electrical finite element modeling for radio frequency cardiac ablation: effects of changes in myocardial properties. *Med Biol Eng Comput* 2000;38:562–568.
28. Haemmerich D, Chachati L, Wright AS, Mahvi DM, Lee Jr FT, Webster JG. Hepatic radiofrequency ablation with internally cooled probes: effect of coolant temperature on lesion size. *IEEE Trans Biomed Eng* 2003;50:493–500.
29. Chang IA, Nguyen UD. Thermal modeling of lesion growth with radiofrequency ablation devices. *Biomed Eng Online* 2004;3:27.
30. Chang IA. Considerations for thermal injury analysis for RF ablation devices. *Biomed Eng Online* 2010;4:3-12.
31. Jacques S, Rastegar S, Thomsen S, Motamedi M. The role of dynamic changes in blood perfusion and optical properties in laser coagulation of tissue. *IEEE J Sel Top Quantum Electron* 1996;2: 922–933.
32. Goldberg SN, Grassi CJ, Cardella JF, Charboneau JW, Dodd 3rd GD, Dupuy DE, et al. Society of Interventional Radiology Technology Assessment Committee; International Working Group on Image-Guided Tumor Ablation. Image-guided tumor ablation: Standardization of terminology and reporting criteria. *J Vasc Interv Radiol* 2009; 20:S377–S390.
33. Dos Santos I, Haemmerich D, Schutt D, Ferreira da Rocha A., Rax Menezes L. Probabilistic finite element analysis of radiofrequency liver ablation using the unscented transform. *Phys Med Biol* 2009; 54(3): 627–640.

Table I. Characteristics of the materials used in the theoretical model [11,14].

Tissue/material	σ (S/m)	k (W/m·K)	ρ (kg/m ³)	c (J/kg·K)
Electrode	7.4×10^6	15	8×10^3	480
Plastic	10^{-5}	0.026	70	1045
Liver	0.132 ⁽¹⁾	0.502 ⁽²⁾	1080 ⁽³⁾	3455 ⁽³⁾
			370 ⁽⁴⁾	2156 ⁽⁴⁾

σ , electric conductivity; k , thermal conductivity; ρ , density; c , specific heat

⁽¹⁾ Assessed at 17°C (initial temperature). This allows initial impedance ($\approx 100 \Omega$) to be matched to the mean value experimentally measured in a previous study.

⁽²⁾ Assessed at 37°C.

⁽³⁾ Normal tissue (liquid phase).

⁽⁴⁾ Gas phase tissue.

Table II. Minor diameter of experimental and theoretical lesions (in cm) created in liver after 720 s of alone RF ablation (RF), and after 720 s of RF ablation combined with previous electroporation (EP+RF). Two thermal damage contours ($\Omega=1$ and $\Omega=4.6$) were used the theoretical case. Not significant differences between EP and RF+EP groups were observed neither in the maximum nor in the minimum diameter.

	EP+RF	RF
Theoretical lesion diameter ($\Omega=1$)	2.98 cm	2.94 cm
Theoretical lesion diameter ($\Omega=4.6$)	2.66 cm	2.63 cm
Experimental lesion diameter (mean \pm standard deviation)	2.0 \pm 0.30 cm (n=3)	1.7 \pm 0.30 cm (n=6)




Figure 1 Geometry of the two-dimensional model used in the theoretical study (out of scale, dimensions in mm). The domain is divided into three zones: plastic portion of the electrode, metallic electrode and hepatic tissue.




Figure 2 Relative blood perfusion after delivery of 8 pulses of $100 \mu\text{s}$ at 1 Hz. The experimental data was obtained from [8]. The model corresponds to expression (2).




Figure 3 Distribution of the relative blood perfusion (α) and the electric field at the end of the application of the EP pulse.




Figure 4 Voltage (A) and impedance (B) evolution in the computer simulations of electroporation + 12-minute RF ablation case without plotting the EP period. (C) Impedance evolution in the computer simulations of alone RF ablation.




Figure 5 Temperature distributions after 12 minutes of RF ablation and previous application of electroporation pulses. The black lines represent the thermal damage lesion which corresponds to $\Omega=1$ and $\Omega=4.6$. The values of the minor transverse lesion diameters (in cm) are shown for both contours of thermal damage.




Figure 6 Temperature distributions after 12 minutes of RF ablation and previous application of EP pulses joint with the blood perfusion level (α) map.




Figure 7 Lesions created in hepatic lobes after: electroporation alone (A), RF ablation alone (B) and electroporation+RF ablation (C and D).